National Institute of Advanced Industrial Science and Technology

National Metrology Institute of Japan



Reference Material Certificate NMIJ CRM 4001-b



No. +++

Ethanol

This certified reference material (CRM) is produced in accordance with the NMIJ's management system and is in compliance with ISO 17034 and ISO/IEC 17025. This CRM is intended for the calibration of instruments, and validation of analytical methods and instruments used for the quantification of ethanol.

Certified Values

The certified values of this CRM are given in the table below. The uncertainty of each certified value is the half-width of the expanded uncertainty interval calculated using a coverage factor (k) of 2, which gives a level of confidence of approximately 95%.

		Certified value,	Expanded uncertainty,	
	CAS No.	Amount-of-substance	Amount-of-substance	
		fraction (mol/mol)	fraction (mol/mol)	
Ethanol	64-17-5	0.999	0.001	
	CAS No.	Certified value,	Expanded uncertainty,	
		Mass fraction (kg/kg)	Mass fraction (kg/kg)	
Ethanol	64-17 <mark>-5</mark>	1.000	0.001	

Analysis

The certified value (amount-of-substance fraction) of this CRM was determined by the freezing point depression method carried out using adiabatic calorimeters operated in accordance with the fractional melting method. The combined standard uncertainty was estimated as a combination of standard uncertainties in purity determination, homogeneity test, and stability test.

The certified value (mass fraction) of this CRM was determined by converting the purity (amount-of-substance fraction) using the estimated average molecular weights of impurities. The combined standard uncertainty was estimated as a combination of standard uncertainties in purity determination, homogeneity test, and stability test.

Metrological Traceability

The certified value (amount-of-substance fraction) was determined by the freezing point depression method carried out using adiabatic calorimeters and are traceable to the International System of Units (SI). Temperature (platinum resistance thermometer), voltage (digital multimeter), resistance (standard resistor), and heating duration (universal counter) of the adiabatic calorimeters were calibrated and were traceable to the SI.

The certified value (mass fraction) was determined by converting the purity (amount-of-substance fraction) using the estimated average molecular weight of impurities. The concentrations of impurities were determined using a gas chromatography with flame ionization detection (GC-FID) and Karl-Fisher (KF) titration. GC-FID was calibrated using solutions prepared gravimetrically with a JCSS-calibrated balance, and the KF method was validated by NMIJ. The certified values were traceable to the SI.

Indicative Values

The concentration of carbon 14 (¹⁴C) in this CRM was determined via the collaborative analysis carried out using accelerator mass spectrometers and liquid scintillation counters. The indicative values of ¹⁴C concentration represented as "percent Modern Carbon" (pMC: Stuiver & Polach (1977) Radiocarbon, 19(3), 355-363) and specific radioactivity are given in the following table. The

quoted uncertainties are the half-width of the expanded uncertainty intervals calculated using the standard deviation of reported results and a coverage factor (k) of 2, which gives a level of confidence of approximately 95 %.

"Specific radioactivity" used in this section means radioactivity per unit mass of carbon. The half-life of ¹⁴C was assumed to be 5730 years (Godwin (1962) Nature, 195, 1984).

	Indicative value	Expanded uncertainty
pMC $(A_{\rm S}/A_{\rm abs} \times 100: \text{ on July } 1, 2011)^{1}$	108.0 %	1.6 %
pMC $(A_{\rm SN}/A_{\rm ON} \times 100)^{2}$	105.8 %	1.6 %
Specific radioactivity (on July 1, 2011) ³⁾	0.2440 Bq/g-carbon	0.0035 Bq/g-carbon

1) A_{abs} denotes the standard "modern" ¹⁴C concentration. It is defined as ¹⁴C concentration in atmospheric CO₂ in 1950, and corresponds to 0.2260 Bq/g-carbon. Practically, NIST SRM 4990C (oxalic acid) is often used as the standard for calculating A_{abs} . In that case, A_{abs} is the specific radioactivity or ¹⁴C/¹²C value of SRM 4990C normalized to –25 ‰ for ¹³C, corrected for decay since 1950, and multiplied by 0.7459. A_S denotes the specific radioactivity or ¹⁴C/¹²C value of this CRM in 2011 (not normalized for ¹³C).

2) $A_{\rm ON}$ denotes the standard ¹⁴C concentration at the time of measurement (in 2011). It is the specific radioactivity or ¹⁴C/¹²C value of SRM 4990C normalized to -25 ‰ for ¹³C, and multiplied by 0.7459. $A_{\rm SN}$ denotes the specific radioactivity or ¹⁴C/¹²C value of this CRM in 2011, normalized to -25 ‰ for ¹³C. The ratio does not depend on the year of measurement.

3) The specific radioactivity of this CRM equals $0.2260 A_s/A_{abs}$. The radioactivity concentration in this CRM (Bq/g-sample) is calculated from specific radioactivity by multiplying it with the carbon content of this CRM (0.5214).

Mutual Recognition Arrangement under Meter Convention

This certificate is consistent with the calibration and measurement capabilities (CMCs) that are included in Appendix C of the Mutual Recognition Arrangement (MRA) drawn up by the International Committee for Weights and Measures (CIPM). Under the MRA, all participating institutes recognize the validity of each other's calibration and measurement certificates for the quantities, ranges and measurement uncertainties specified in Appendix C (as for Appendix C of MRA, see https://www.bipm.org/kcdb/).

Expiration of Certification

This certificate is valid for one year from the date of shipment, provided that this CRM is stored in accordance with the instructions given in this certificate.

Description of the material

This CRM is a high purity ethanol. This CRM is in the form of a colorless and clear liquid at room temperature and 15 mL in net volume is sealed in an amber glass ampoule containing argon gas.

Instructions for Storage

This CRM should be stored in a cold (temperature: about -20 °C), clean and dark place.

Instructions for Use

This CRM is for laboratory use only. The ampoule of this CRM should be allowed to warm to room temperature and then should be opened at the dew point as low as possible. The CRM should be used promptly once the ampoule is opened.

Precautions for Handling

Keep away from heat and ignition sources. The CRM should be used in a well-ventilated place. Wear personal protective equipment such as safety glasses, safety mask, and safety gloves in handling. Refer to the safety data sheet (SDS) on this CRM before use.

Preparation

This CRM was purified and subdivided by Kanto Chemical Co., Inc. It was purified by distillation and drying followed by addition

of ultrapure water to be approximately 100 mg/kg. Fifteen mL of ethanol was filled into an amber glass ampoule in an argon atmosphere.

Technical Information

At the characterization, the concentrations of water and acetaldehyde in this CRM were 108 mg/kg and 2 mg/kg, respectively. These impurities were measured by KF and GC-FID, respectively. The δ^{13} C value determined by isotope–ratio mass spectrometry was –11.7 ‰ (standard deviation: 0.1 ‰, n = 3).

Collaborators

Measurements for the ¹⁴C content and δ^{13} C value were conducted as a collaborative analysis among 13 laboratories (listed below) coordinated by the NMIJ.

Institute of Accelerator Analysis, Ltd. The General Environmental Technos, Co., Ltd. Kyushu Environmental Evaluation Association Geo Science Laboratory, Co., Ltd. Tokyo Metropolitan Industrial Technology Research Institute Radioisotope Center, The University of Tokyo Atmosphere and Ocean Research Institute, The University of Tokyo Center for Chronological Research, Nagoya University Japan Radioisotope Association Nippon Kaiji Kentei Kyokai Japan Chemical Analysis Center Paleo Labo, Co., Ltd. Quantum Radiation Division, National Metrology Institute of Japan

NMIJ Analysts

For this CRM, the technical manager is KATO K., the person responsible for production is KITAMAKI Y., and the production analysts are KITAMAKI Y., SHIMIZU Y., INAGAKI S., YAMAZAKI T., NUMATA M. and YOSHIMURA E.

Information

If substantive technical changes occur that affect the certification before the expiration of this certificate, NMIJ will notify the registered customer. Customer registration on the NMIJ Website (given below) will facilitate notification. Technical reports regarding this CRM can be obtained from the contact details given below.

Reproduction of Certificate

In reproducing this certificate, it should be clearly indicated that the document is a copy.

April 1, 2020

ISHIMURA Kazuhiko President National Institute of Advanced Industrial Science and Technology

If you have any questions about this CRM, please contact: National Institute of Advanced Industrial Science and Technology, National Metrology Institute of Japan, Center for Quality Management of Metrology, Reference Materials Office, 1-1-1 Umezono, Tsukuba, Ibaraki 305-8563, Japan Phone: +81-29-861-4059; Fax: +81-29-861-4009, https://unit.aist.go.jp/nmij/english/refmate/

Revision history

April 1, 2015: "Metrology Management Center" was renamed to "Center for Quality Management of Metrology." December 24, 2020: The description in "Expiration of Certification" was changed to "one year from the date of shipment."

