

National Institute of Advanced Industrial Science and Technology

National Metrology Institute of Japan



Reference Material Certificate

NMIJ CRM 4048-a

No. +++

Simazine



This certified reference material (CRM) was produced based on NMIJ's quality system in compliance with JIS Q 0034 (ISO GUIDE 34). Intended uses for this CRM are calibration of analytical instruments and reference materials, quality control of analytical instrument and validation of analytical techniques and instruments.

Certified Value

The certified value for the purity (in mass fraction) of simazine is given in the following table. The expanded uncertainty was determined using coverage factor $k=2$, corresponding to an estimated confidence interval of approximately 95%.

	CAS Number	Certified Value Mass fraction (kg/kg)	Expanded Uncertainty Mass fraction (kg/kg)
Simazine (6-chloro-N,N'-diethyl-1,3,5-triazine-2,4-diamine)	122-34-9	0.996	0.002

Analytical Methods

The certified value was determined by subtracting the mass fractions of impurities measured by gas-chromatography with flame ionization detection (GC-FID), high performance liquid chromatography (HPLC) and Karl-Fischer titration (KF).

Traceability

The certified value of this CRM was determined by subtraction method. Mass fractions of major impurities including ethyl acetate and atrazine were determined by using their calibration standards evaluated at NMIJ. The mass fraction of simazine was obtained by subtracting the mass fractions of the impurities from 1. It is traceable to the International System of Units (SI).

Expiration of Certification

The certification of this CRM is valid until March 31, 2020, provided that the material is unopened and stored in accordance with the instructions given in this certificate.

Sample Form

The form of this CRM is white powder and a net amount of 100 mg is enclosed in a 10 mL amber glass vial, and packed in an aluminum-laminated bag.

Homogeneity

The homogeneity of the CRM was determined by analyzing ten vials randomly selected out of 318 subdivided vials. Area percentages of simazine measured by HPLC were evaluated to prove the homogeneity of the CRM. Variation of purity (mass fraction) between vials was estimated by the analysis of variance and was reflected in the uncertainty estimation of the certified value.

Precautions for Storage

This CRM should be stored in a refrigerator (around 5 °C) at a clean and dark place.

[This document is just explanation translated from the original Japanese certificate and some information is omitted from it.]

Precautions for use

Sampling more than 10 mg is recommended for ensuring homogeneity. Keep the CRM at room temperature before opening and use quickly after opening. Do not use the CRM other than testing and research purposes. The use, handling and storage of this CRM should be made observing the laws regulating those of simazine.

Precautions for handling

Wear protective equipments such as a safety mask and safety gloves in handling.

Preparation Method

Raw material of this CRM was commercially available simazine pre-selected by the quality evaluation by NMIJ. This CRM was prepared and subdivided into vials by Wako Pure Chemical Industries, Ltd. Vials were packed in an aluminum-laminated bags under Ar gas atmosphere.

Information Values

This CRM contains ethyl acetate (2.1 g/kg, mass fraction) and atrazine (0.3 g/kg, mass fraction).

NMIJ Analysts

The technical manager for this CRM is M. Numata. The production manager is K. Ishikawa. The analysts are K. Ishikawa, N. Hanari, T. Saito, S. Otsuka, K. Higuchi, N. Fujiki, R. Iwasawa and T. Miura.

Technical Information

Customers will be notified of any revision to this CRM including a change in certified value. Technical information on this CRM can be obtained from the home page and other contact shown below.

Reproduction of Certificate

In reproducing this certificate, it should be clearly indicated that the document is a copy.

May 19, 2010

Tamotsu Nomakuchi

President

National Institute of Advanced Industrial Science and Technology

If you have any questions about this CRM, please contact
National Institute of Advanced Industrial Science and Technology,
National Metrology Institute of Japan,
Metrology Management Center, Reference Materials Office,
1-1-1, Umezono, Tsukuba, Ibaraki 305-8563, Japan
Phone: +81-29-861-4059; Fax: +81-29-861-4009, <http://www.nmij.jp/>

Note: This certificate is a translation of the original Japanese certificate and is not an official document.