

National Institute of Advanced Industrial Science and Technology

National Metrology Institute of Japan



Reference Material Certificate

NMIJ CRM 7603-a
No. +++

Seawater for Nutrients - High Concentration

This certified reference material (CRM) is produced in accordance with the NMIJ's management system and is in compliance with ISO 17034 and ISO/IEC 17025. This CRM is intended for use in controlling the precision of analysis and validating analytical methods and instruments during analysis of nutrients in seawater.

Certified Values

The certified values of this CRM are given in the table below. The uncertainty of each certified value is the expanded uncertainty obtained by multiplying the combined standard uncertainty by a coverage factor (k) of 2, and it is the half-width of an interval estimated to have a level of confidence of approximately 95 %.

Substance	Certified value, *1 Mass fraction (mg/kg)	Expanded uncertainty, Mass fraction (mg/kg)	Analytical methods *2
Nitrate ions	2.745	0.050	1), 3), 4)
Dissolved silica	4.106	0.043	1), 2), 5)

*1 The certified value for dissolved silica is expressed as the mass fraction of silicon.

*2 Analytical methods: See the next section "Analysis".

Analysis

The certified values of this CRM were arithmetic mean of the results of the following analytical methods:

- 1) Colorimetric method (continuous flow mode and batch one)
- 2) Ion exclusion chromatography/isotope dilution-inductively coupled plasma mass spectrometry
- 3) Ion chromatography by direct analysis
- 4) Ion chromatography after halogen-ion separation
- 5) Ion exclusion chromatography with post-column detection

Metrological Traceability

Each certified value was determined by more than one method with using one of NIST (National Institute of Standards and Technology) SRM 3150 of a silicon standard solution and an NMIJ primary standard solution of nitrate ions. The certified values, therefore, are traceable to the International System of Units (SI).

Indicative Value

The indicative value was determined with the NMIJ primary standard solution of phosphate ions. The indicative value of this CRM is given in the table below.

Substance	Indicative value, Mass fraction (mg/kg)	Analytical method (see the section "Analysis" (<i>vide supra</i>))
Phosphate ions	0.288 ± 0.005	1)

The value following the \pm in the column of the indicative value is the half-width of the expanded uncertainty interval calculated using a coverage factor (k) of 2, which gives a level of confidence of approximately 95 %.

Expiration of Certification

This certificate is valid for one year from the date of shipment, provided that this CRM remains unopened and is stored in accordance with the instructions given in this certificate.

Description of the material

This CRM is seawater collected from the nutrient maximum layer in the Pacific Ocean. This CRM is in the form of a colorless and transparent liquid at ordinary temperature, and *ca.* 90 mL is kept in a polypropylene bottle. The bottle is sealed in an aluminum-laminated plastic bag.

Instructions for Storage

This CRM should be kept in the polypropylene bottle sealed in the aluminum-laminated plastic bag. This CRM should be stored in a clean place at temperatures of 5 °C to 30 °C and protected from light.

Instructions for Use

The bottle should be allowed to warm to room temperature before opening. Prior to use, the bottle should be shaken thoroughly but gently at room temperature. This CRM is for laboratory use only.

Precautions for Handling

Refer to the safety data sheet (SDS) on this CRM before use.

Preparation

The raw material seawater of this CRM was collected from the nutrient maximum layer in the Pacific Ocean with the cooperation of the Meteorological Research Institute, Japan Meteorological Agency (MRI) and the Japan Agency for Marine-Earth Science and Technology (JAMSTEC). The seawater was sterilized in an autoclave of stainless steel, followed by subdividing in polypropylene bottles and sealing by the aluminum-laminated plastic bags in a clean room. This bottling process was carried out with cooperation of the General Environmental Technos Co., Ltd. (current: KANSO Co., Ltd.).

Technical Information

The value 1.023 g/cm³ (25 °C) can be used as the density of the seawater for the purpose of air-buoyancy correction. This CRM contains a silicon compound which cannot be detected by the colorimetric method using the molybdenum blue method or the molybdenum yellow method, but can be separated and detected by ion exclusion chromatography-inductively coupled plasma mass spectrometry. This CRM contains nitrite ions (0.0012 mg/kg), which was determined with the NMIJ primary standard solution of nitrite ions. Concentration of nitrite ions in this CRM were measured by colorimetric method (continuous flow mode) at August 2016.

NMIJ Analysts

The technical manager for this CRM is HIOKI A., the production manager is MIURA T. and the analysts are CHEONG C., NONOSE N., SUZUKI T., MIURA T., YAMAUCHI Y., and ISHIZAWA Y.

Information

If substantive technical changes occur that affect the certification before the expiration of this certificate, NMIJ will notify the registered customer. Customer registration on the NMIJ Website (given below) will facilitate notification. Technical reports regarding this CRM can be obtained from the contact details given below.

Reproduction of Certificate

In reproducing this certificate, it should be clearly indicated that the document is a copy.

April 1, 2020

ISHIMURA Kazuhiko

President

National Institute of Advanced Industrial Science and Technology

If you have any questions about this CRM, please contact
National Institute of Advanced Industrial Science and Technology,
National Metrology Institute of Japan,
Center for Quality Management of Metrology, Reference Materials Office,
1-1-1 Umezono, Tsukuba, Ibaraki 305-8563, Japan
Phone: +81-29-861-4059; Fax: +81-29-861-4009, <https://unit.aist.go.jp/nmij/english/refmate/>

Revision history

- April 1, 2015: "Metrology Management Center" was renamed to "Center for Quality Management of Metrology."
- September 6, 2016: The description in "Expiration of Certification" was changed to "one year from the date of shipment."
Certified values of nitrate ions and dissolved silica were revised. Indicative value of phosphate ions was revised. Indicative value of nitrite ions was changed to information.
- August 24, 2018: "Indicative Value" the standard solution used for determination of phosphate ions was described.
"Information" the standard solution used for determination of nitrite ions was described.
- January 21, 2021: "Sample Form" was changed to "Description of the material" and collection site of the seawater was added.
"Certified values" the uncertainty of nitrate ion and dissolved silica were revised.