

National Institute of Advanced Industrial Science and Technology

National Metrology Institute of Japan



Reference Material Certificate

NMIJ CRM 8133-a
No. +++)

Heavy Metals (Cd, Cr, Hg, Pb) in PP Resin Pellet- High Concentration

This certified reference material (CRM) is produced in accordance with the NMIJ's management system and is in compliance with ISO 17034 and ISO/IEC 17025. This CRM is intended for use in controlling the precision of analysis and for validating analytical methods and instruments during the quantitative determination of Cd, Cr, Hg and Pb in polypropylene resin and similar polymers.

Certified Values

The certified values of Cd, Cr, Hg and Pb in this CRM are given in the table below. The drying instruction is described in this certificate. The uncertainty of the certified value is the half-width of the expanded uncertainty interval calculated using a coverage factor (k) of 2, which gives a level of confidence of approximately 95 %.

	Certified value, Mass fraction (mg/kg)	Expanded uncertainty, Mass fraction (mg/kg)
Cd	94.26	1.39
Cr	895.2	9.6
Hg	941.5	19.6
Pb	949.2	7.5

Analysis

Each certified value of this CRM was determined by the following analytical methods:

- (1) Microwave digestion using sulfuric acid and nitric acid / isotope dilution – inductively coupled plasma mass spectrometry (Cd, Cr, Hg and Pb),
- (2) Dry-ashing digestion followed by open-system dissolution using nitric acid / inductively coupled plasma atomic emission spectrometry (Cd and Pb),
- (3) Microwave digestion using 70% nitric acid / inductively coupled plasma atomic emission spectrometry (Cd, Cr, Hg and Pb).

Metrological Traceability

Each certified value of this CRM was determined by more than one method, including isotope dilution – mass spectrometry as the primary method of measurement with NMIJ primary standard solutions of Cd, Cr, Hg and Pb. The certified values are traceable to the International System of Units (SI).

Mutual Recognition Arrangement under Meter Convention

This certificate is consistent with the calibration and measurement capabilities (CMCs) that are included in Appendix C of the Mutual Recognition Arrangement (MRA) drawn up by the International Committee for Weights and Measures (CIPM). Under the MRA, all participating institutes recognize the validity of each other's calibration and measurement certificates for the quantities, ranges and measurement uncertainties specified in Appendix C (as for Appendix C of MRA, see <http://kcdb.bipm.org/AppendixC/default.asp>).

Expiration of Certification

This certificate is valid until March 31, 2021, provided that the material remains unopened and is stored in accordance with the instructions given in this certificate.

Sample Form

This CRM is in the form of a small pellets and 25 g in net mass is kept in a brown glass bottle.

Homogeneity

The homogeneity of this CRM was determined by analyzing 14 bottles selected at approximately equal intervals on the basis of the order of bottling. The elements (Cd, Cr, Hg and Pb) were determined by microwave digestion using 70% nitric acid / inductively coupled plasma mass spectrometry (using an external calibration curve and an internal standard). The homogeneity of each element is reflected in the uncertainty of the certified value.

Instructions for Storage

This CRM should be stored at a temperature between 15 °C and 35 °C, and shielded from light.

Instructions for Use

Prior to use, the sample should be mixed by slow rolling. It should be dried for one hour at 80 °C and then maintained at room temperature for one hour in a silica-gel desiccator. The recommended sample mass is 0.10 g or more for one analysis. Please note that this sample contains tens mg/kg of Br.

Precautions for Handling

Do not use this standard substance for testing/research purposes only. Pay attention to fire and ventilation; wear protective mask, protective gloves, etc. DBDE (deca brominated biphenyl ether) is designated as Class 1 Designated Chemical Substances in the Act on the Evaluation of Chemical Substances and Regulation of Their Manufacture, etc. It is also designated as Class I Designated Chemical Substances in the Act on Confirmation, etc. of Release Amounts of Specific Chemical Substances in the Environment and Promotion of Improvements to the Management Thereof (PRTR law). Handle in compliance with these laws. Refer to the safety data sheet (SDS) on this CRM before use.

Preparation

The base resin was prepared by mixing a commercial PP resin and powders of CdO, PbCrO₄, chromium(III) acetylacetonate, and HgS were mixed and then pellets were produced from the mixture. The pellet-producing process was repeated two more times.

NMIJ Analysts

The technical and production managers are HIOKI A. and the analysts are OHATA M. and HIOKI A.

Information

If substantive technical changes occur that affect the certification before the expiration of this certificate, NMIJ will notify the registered customer. Customer registration on the NMIJ Website (given below) will facilitate notification. Technical reports regarding this CRM can be obtained from the contact details given below.

Reproduction of Certificate

In reproducing this certificate, it should be clearly indicated that the document is a copy.

April 1, 2020

ISHIMURA Kazuhiko
President

National Institute of Advanced Industrial Science and Technology

If you have any questions about this CRM, please contact
National Institute of Advanced Industrial Science and Technology,

Revision history

August 10, 2012: The expiration of certification was extended from March 31, 2013 to March 31, 2021.

April 1, 2015: "Metrology Management Center" was renamed to "Center for Quality Management of Metrology."

National Metrology Institute of Japan,
Center for Quality Management of Metrology, Reference Materials Office,
1-1-1 Umezono, Tsukuba, Ibaraki 305-8563, Japan
Phone: +81-29-861-4059, Fax: +81-29-861-4009, <https://unit.aist.go.jp/nmij/english/refmate/>

SAMPLE